

# EU DECLARATION OF CONFORMITY

according to Medical Device regulation (EU) 2017/745 and  
according to Low Voltage directive 2014/35/EU and  
according to EMC directive 2014/30/EU and  
according to RoHS directive 2011/65/EU and  
according to REACH directive 1907/2006/EU and  
according to Eco-Design framework directive 2009/125/EC and  
according to WEEE directive 2012/19/EU

We as manufacturer: **flow-med GmbH, Weinbergweg 10, 73773 Aichwald, Germany**

SRN: **DE-MF-000006336**

declare under our responsibility that the following listed medical devices meet all applicable basic safety and performance requirements of the regulations mentioned above.

Product name: **vac-clean uv 2®**  
Item number: **621 00000**  
Basic UDI-DI: **4260277010076R**  
EMDN: **Y214580**  
Intended use: **Electronic vacuum cleaner for hearing aids**  
Product classification: **Class 1, according to annex VIII, rule 1 and 13 of regulation (EU) 2017/745 (MDR)**  
Applied specifications/ norms: **EN 55014:2021 and EN63000:2018**



Valid until:  
**31.12.2025**

**Aichwald, 01.01.2024**

Place, date

**CE**

**MD**

Name Managing Director:  
**Meiko Panitzsch**

**Flow** med  
medical appliances

Signature **Weinbergweg 10**  
**73773 Aichwald**  
**Germany**

phone **+49 711 50463500**  
fax **+49 711 50463501**